

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3987

Name: RITTER2  
Sterilizer ID: V2313228  
10/02/2024 10:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.6 F 0.0  
2:00 132.3 F 0.7  
4:00 144.5 F 2.8  
6:00 163.4 F 5.7  
8:00 181.9 F 9.2  
10:00 204.5 F 13.0  
12:00 216.9 F 6.5  
14:00 228.9 F 10.4  
16:00 229.2 F 8.6  
18:00 238.5 F 11.0  
20:00 234.0 F 8.4  
22:00 247.2 F 14.2  
24:00 257.0 F 19.7  
26:00 265.7 F 25.5  
28:00 272.9 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.6  
1:00 271.7 F 29.5  
2:00 272.1 F 29.5  
3:00 272.4 F 29.6  
4:00 272.1 F 29.8

Min 271.5 F 29.5  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.9 F 0.0  
5:00 207.8 F 0.0  
10:00 188.5 F 0.0  
15:00 185.0 F 0.0  
20:00 182.3 F 0.0  
25:00 181.7 F 0.0  
30:00 179.9 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 28:22  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:06:58  
CYCLE COMPLETE

10/02/2024 11:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_