
Midmark M11 - v1.0.5
Total Cycles: 1290

Name: RITTER2
Sterilizer ID: V2313228
07/13/2022 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.3 F 0.0
2:00 149.5 F 0.0
4:00 172.3 F 0.5
6:00 200.9 F 4.3
8:00 217.7 F 8.8
10:00 230.7 F 11.3
12:00 233.3 F 10.3
14:00 233.3 F 8.5
16:00 242.4 F 12.2
18:00 244.8 F 13.0
20:00 256.9 F 19.5
22:00 268.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.3 F 29.5
2:00 271.3 F 28.8
3:00 272.2 F 29.6
4:00 271.6 F 28.9

Min 271.3 F 28.7
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 215.9 F 0.0
10:00 227.9 F 0.0
15:00 226.5 F 0.0
20:00 224.4 F 0.0
25:00 222.4 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:19
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:30
CYCLE COMPLETE

07/13/2022 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____