
Midmark M11 - v1.0.5
Total Cycles: 1406

Name: RITTER2
Sterilizer ID: V2313228
08/24/2022 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.7 F 0.0
2:00 155.1 F 1.5
4:00 170.7 F 4.0
6:00 194.3 F 8.5
8:00 214.3 F 11.8
10:00 224.9 F 8.1
12:00 236.9 F 13.1
14:00 237.5 F 10.6
16:00 235.7 F 8.9
18:00 249.7 F 15.5
20:00 261.0 F 22.0
22:00 271.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.8
2:00 271.9 F 29.1
3:00 271.4 F 29.0
4:00 272.4 F 29.8

Min 271.3 F 28.7
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 209.2 F 0.0
10:00 237.2 F 0.0
15:00 236.2 F 0.0
20:00 224.2 F 0.0
25:00 220.8 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

08/24/2022 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____