
Midmark M11 - v1.0.5
Total Cycles: 2714

Name: RITTER2
Sterilizer ID: V2313228
07/31/2023 04:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.1 F 0.0
2:00 134.5 F 0.0
4:00 153.7 F 0.0
6:00 186.1 F 0.1
8:00 207.5 F 4.3
10:00 221.3 F 8.6
12:00 232.3 F 13.2
14:00 234.4 F 9.9
16:00 236.1 F 9.2
18:00 242.9 F 11.8
20:00 247.0 F 13.8
22:00 259.5 F 20.7
24:00 270.6 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.0
2:00 271.9 F 29.0
3:00 272.3 F 28.8
4:00 272.0 F 29.1

Min 271.4 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 208.5 F 0.0
10:00 213.0 F 0.0
15:00 214.9 F 0.0
20:00 214.8 F 0.0
25:00 215.3 F 0.0
30:00 213.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:46
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:59
CYCLE COMPLETE

07/31/2023 05:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____