
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001235
10:51 30/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:28
STERILIZING 3:32
136.4 C 223kPa 3:32
136.3 C 224kPa 4:02
136.3 C 224kPa 4:32
136.3 C 224kPa 5:02
136.3 C 224kPa 5:32
136.4 C 223kPa 6:02
135.9 C 222kPa 6:32
136.1 C 222kPa 7:02
136.1 C 222kPa 7:02
VENTING 7:02
AIR DRYING 7:50
CYCLE COMPLETE 22:50

Digital Signature #
770F49E9385456CB
