
Midmark M11 - v1.0.0
Total Cycles: 801

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2021 12:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.3 F 0.0
2:00 135.0 F 0.9
4:00 165.7 F 2.7
6:00 187.2 F 6.5
8:00 206.1 F 11.1
10:00 214.8 F 4.0
12:00 226.9 F 8.8
14:00 238.3 F 12.7
16:00 238.1 F 10.6
18:00 235.3 F 8.4
20:00 249.3 F 15.1
22:00 261.0 F 21.6
24:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.4
3:00 272.6 F 29.2
4:00 271.8 F 28.5

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 210.1 F 0.0
10:00 223.1 F 0.0
15:00 225.0 F 0.0
20:00 226.4 F 0.0
25:00 227.9 F 0.0
30:00 242.2 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:36
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

04/22/2021 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____