

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1091

Name: RITTER2  
Sterilizer ID: V2313228  
05/11/2022 07:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 75.0 F 0.0  
4:00 86.1 F 0.0  
6:00 122.5 F 0.0  
8:00 156.6 F 0.2  
10:00 185.8 F 1.4  
12:00 200.0 F 5.3  
14:00 212.4 F 8.9  
16:00 224.8 F 12.9  
18:00 228.9 F 7.9  
20:00 240.5 F 12.7  
22:00 239.3 F 10.5  
24:00 236.7 F 8.9  
26:00 251.1 F 15.7  
28:00 262.7 F 22.4  
30:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.7 F 28.9  
2:00 271.6 F 28.5  
3:00 272.6 F 28.9  
4:00 271.5 F 28.4

Min 271.5 F 28.2  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 171.2 F 0.0  
5:00 185.2 F 0.0  
10:00 197.0 F 0.0  
15:00 194.2 F 0.0  
20:00 192.9 F 0.0  
25:00 190.0 F 0.0  
30:00 199.1 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 30:17  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 01:08:10  
CYCLE COMPLETE

05/11/2022 08:38 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_