
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000629
10:02 03/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:55
STERILIZING 5:06
136.0 C 229kPa 5:06
136.3 C 232kPa 5:37
136.3 C 232kPa 6:07
136.3 C 230kPa 6:37
136.3 C 231kPa 7:07
136.4 C 232kPa 7:37
136.3 C 231kPa 8:07
136.4 C 230kPa 8:37
136.4 C 230kPa 8:37
VENTING 8:37
AIR DRYING 9:31
CYCLE COMPLETE 24:31

Digital Signature #
42B25E4333906860
