
Midmark M11 - v1.0.0
Total Cycles: 4341

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2024 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.9 F 0.0
2:00 151.2 F 1.2
4:00 148.5 F 3.6
6:00 157.2 F 6.6
8:00 197.1 F 10.4
10:00 219.8 F 4.8
12:00 216.4 F 7.2
14:00 226.9 F 11.5
16:00 229.8 F 7.6
18:00 240.3 F 13.0
20:00 240.6 F 11.2
22:00 252.2 F 16.9
24:00 262.8 F 23.0
26:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.4 F 29.2
2:00 271.8 F 29.4
3:00 272.0 F 29.0
4:00 272.1 F 29.6

Min 271.2 F 28.8
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 215.3 F 0.0
10:00 220.3 F 0.0
15:00 223.9 F 0.0
20:00 236.1 F 0.0
25:00 231.4 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:25
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:05:12
CYCLE COMPLETE

02/28/2024 12:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____