

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3597

Name: RITTER 1  
Sterilizer ID: V2178198  
08/01/2023 07:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.7 F 0.0  
2:00 77.8 F 0.0  
4:00 88.0 F 0.0  
6:00 113.6 F 0.0  
8:00 147.9 F 0.1  
10:00 181.2 F 0.4  
12:00 200.8 F 1.8  
14:00 214.4 F 4.7  
16:00 219.5 F 7.5  
18:00 228.9 F 10.7  
20:00 233.7 F 7.9  
22:00 237.9 F 10.3  
24:00 238.9 F 8.7  
26:00 246.1 F 13.2  
28:00 243.6 F 11.8  
30:00 255.4 F 17.9  
32:00 265.9 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.3 F 28.3  
2:00 272.4 F 28.7  
3:00 271.8 F 28.0  
4:00 272.7 F 28.7  
  
Min 271.5 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
5:00 204.5 F 0.0  
10:00 208.0 F 0.0  
15:00 208.0 F 0.0  
20:00 209.7 F 0.0  
25:00 212.9 F 0.0  
30:00 214.8 F 0.0  
  
Drying Complete

FILLING: 0:53  
HEATING: 33:45  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:12:01  
CYCLE COMPLETE

08/01/2023 09:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_