
Midmark M11 - v1.0.5
Total Cycles: 3700

Name: RITTER2
Sterilizer ID: V2313228
06/14/2024 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.3 F 0.0
2:00 168.7 F 2.2
4:00 176.5 F 6.1
6:00 200.0 F 10.9
8:00 217.1 F 5.4
10:00 229.3 F 10.7
12:00 230.8 F 7.5
14:00 241.2 F 12.1
16:00 244.2 F 13.0
18:00 256.6 F 19.6
20:00 267.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.4 F 29.6
2:00 271.8 F 29.8
3:00 271.9 F 29.4
4:00 272.1 F 30.0

Min 271.2 F 29.2
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 207.7 F 0.0
10:00 208.9 F 0.0
15:00 212.3 F 0.0
20:00 211.9 F 0.0
25:00 208.2 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

06/14/2024 10:39 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____