

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002051  
8:38 18/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:58  
STERILIZING 2:54  
136.0 C 220kPa 2:54  
136.4 C 226kPa 3:24  
136.4 C 222kPa 3:54  
136.4 C 226kPa 4:24  
136.4 C 226kPa 4:54  
136.4 C 224kPa 5:24  
136.3 C 224kPa 5:54  
136.2 C 223kPa 6:24  
136.2 C 223kPa 6:24  
VENTING 6:24  
AIR DRYING 7:14  
CYCLE COMPLETE 37:14  
-----

Digital Signature #  
A5F3F7552AD98365  
-----