
Midmark M11 - v1.0.5
Total Cycles: 438

Name: RITTER2
Sterilizer ID: V2313228
10/12/2021 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.2 F 0.0
2:00 141.4 F 1.4
4:00 146.0 F 4.6
6:00 170.3 F 8.4
8:00 192.6 F 12.7
10:00 214.6 F 6.8
12:00 224.2 F 11.6
14:00 233.0 F 8.2
16:00 245.0 F 13.3
18:00 243.6 F 12.1
20:00 255.2 F 18.1
22:00 265.4 F 24.5
24:00 271.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.4 F 29.5
2:00 272.5 F 29.2
3:00 271.8 F 29.5
4:00 272.2 F 29.6

Min 271.4 F 29.1
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 243.8 F 0.0
10:00 228.2 F 0.0
15:00 229.5 F 0.0
20:00 228.6 F 0.0
25:00 226.7 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:25
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:02:50
CYCLE COMPLETE

10/12/2021 11:37 AM
ACCEPTED: X
REJECTED:
Operator: U01