

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3241

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2023 05:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.6 F 0.0  
2:00 180.6 F 3.6  
4:00 202.6 F 8.6  
6:00 218.3 F 12.1  
8:00 225.6 F 9.0  
10:00 235.8 F 10.3  
12:00 242.3 F 12.7  
14:00 243.5 F 12.0  
16:00 256.1 F 18.7  
18:00 267.2 F 25.7  
20:00 271.9 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 273.0 F 29.6  
2:00 271.4 F 28.8  
3:00 272.9 F 29.5  
4:00 271.4 F 28.8  
  
Min 271.2 F 28.8  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
5:00 235.3 F 0.0  
10:00 223.7 F 0.0  
15:00 228.3 F 0.0  
20:00 231.6 F 0.0  
25:00 228.6 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 20:01  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 00:59:33  
CYCLE COMPLETE

12/20/2023 06:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_