
Midmark M11 - v1.0.5
Total Cycles: 2661

Name: RITTER2
Sterilizer ID: V2313228
07/18/2023 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.5 F 0.0
2:00 193.3 F 3.1
4:00 196.5 F 7.7
6:00 214.7 F 12.7
8:00 223.7 F 7.7
10:00 235.8 F 12.7
12:00 237.1 F 10.4
14:00 235.8 F 8.6
16:00 248.7 F 15.1
18:00 259.7 F 21.2
20:00 269.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.8
2:00 272.5 F 29.5
3:00 271.8 F 29.0
4:00 271.6 F 29.2

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 215.9 F 0.0
10:00 224.2 F 0.0
15:00 212.5 F 0.0
20:00 210.3 F 0.0
25:00 210.1 F 0.0
30:00 210.1 F 0.0
Drying Complete

FILLING: 2:21
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

07/18/2023 11:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____