
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000163
13:52 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:42
STERILIZING 5:16
136.0 C 220kPa 5:16
136.2 C 224kPa 5:46
136.4 C 225kPa 6:16
136.2 C 224kPa 6:46
136.4 C 226kPa 7:16
136.3 C 224kPa 7:46
136.3 C 224kPa 8:16
136.3 C 224kPa 8:46
136.3 C 224kPa 8:46
VENTING 8:47
AIR DRYING 9:48
CYCLE COMPLETE 9:55

Digital Signature #
1C54778D854E2169
