

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4643

Name: RITTER 1  
Sterilizer ID: V2178198  
06/14/2024 11:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.8 F 0.0  
2:00 132.8 F 1.9  
4:00 156.6 F 4.8  
6:00 188.7 F 8.3  
8:00 209.9 F 12.5  
10:00 221.6 F 5.6  
12:00 230.0 F 9.7  
14:00 237.5 F 9.7  
16:00 239.6 F 10.7  
18:00 235.0 F 7.9  
20:00 247.9 F 14.3  
22:00 259.1 F 20.4  
24:00 268.8 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 271.5 F 28.8  
2:00 271.4 F 28.8  
3:00 271.5 F 28.7  
4:00 271.9 F 28.7

Min 271.4 F 28.7  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.3 F 0.0  
5:00 188.4 F 0.0  
10:00 188.7 F 0.0  
15:00 189.5 F 0.0  
20:00 190.0 F 0.0  
25:00 190.1 F 0.0  
30:00 189.7 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 25:36  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:04:13  
CYCLE COMPLETE

06/14/2024 12:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_