
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001849
12:30 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:46
STERILIZING 3:53
136.0 C 221kPa 3:53
136.4 C 224kPa 4:23
136.4 C 226kPa 4:53
136.4 C 224kPa 5:23
136.2 C 224kPa 5:53
136.1 C 223kPa 6:23
136.1 C 222kPa 6:53
136.4 C 224kPa 7:23
136.4 C 224kPa 7:23
VENTING 7:24
AIR DRYING 8:16
CYCLE COMPLETE 11:06

Digital Signature #
1629A21E7EE181E2
