
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002365
12:02 06/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:59
STERILIZING 3:08
136.5 C 226kPa 3:08
136.2 C 223kPa 3:38
136.4 C 225kPa 4:08
136.3 C 223kPa 4:38
136.4 C 226kPa 5:08
136.3 C 224kPa 5:38
136.4 C 225kPa 6:08
136.3 C 224kPa 6:38
136.3 C 224kPa 6:38
VENTING 6:38
AIR DRYING 7:28
CYCLE COMPLETE 14:50

Digital Signature #
56FBBF65678B886B
