

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001570  
14:05 11/10/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:38  
PRESSURIZING 6:24  
STERILIZING 7:59  
136.0 C 220kPa 7:59  
136.2 C 222kPa 8:29  
136.2 C 221kPa 8:59  
136.2 C 221kPa 9:29  
136.4 C 223kPa 9:59  
136.4 C 222kPa 10:29  
136.5 C 220kPa 10:59  
136.4 C 223kPa 11:29  
VENTING 11:29  
AIR DRYING 12:32  
CYCLE COMPLETE 18:57  
-----

Digital Signature #  
F964B5AE12FE2C06  
-----