
Midmark M11 - v1.0.0
Total Cycles: 2396

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2022 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 167.1 F 2.6
4:00 180.3 F 6.2
6:00 202.6 F 10.7
8:00 218.6 F 4.8
10:00 230.1 F 10.3
12:00 231.2 F 7.5
14:00 242.3 F 12.6
16:00 240.8 F 10.8
18:00 253.4 F 17.2
20:00 264.6 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.1 F 28.9
2:00 271.7 F 28.9
3:00 272.0 F 28.8
4:00 271.7 F 28.9

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 225.6 F 0.0
10:00 245.6 F 0.0
15:00 235.3 F 0.0
20:00 225.4 F 0.0
25:00 224.6 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

10/04/2022 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____