
Midmark M11 - v1.0.0
Total Cycles: 3510

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2023 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.2 F 0.0
2:00 151.2 F 0.0
4:00 156.4 F 1.7
6:00 182.2 F 4.8
8:00 205.7 F 9.0
10:00 221.0 F 9.9
12:00 227.5 F 8.2
14:00 239.1 F 13.1
16:00 237.9 F 10.2
18:00 235.7 F 8.2
20:00 248.5 F 14.6
22:00 260.5 F 21.2
24:00 271.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.0
2:00 271.6 F 28.5
3:00 272.5 F 29.0
4:00 271.5 F 28.5

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 209.1 F 0.0
10:00 215.4 F 0.0
15:00 224.9 F 0.0
20:00 232.1 F 0.0
25:00 220.9 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:43
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

07/11/2023 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____