
Midmark M11 - v1.0.5
Total Cycles: 713

Name: RITTER2
Sterilizer ID: V2313228
01/20/2022 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.8 F 0.0
2:00 189.5 F 4.4
4:00 209.9 F 10.3
6:00 222.0 F 6.1
8:00 236.8 F 12.2
10:00 238.8 F 10.5
12:00 238.5 F 9.9
14:00 253.6 F 17.3
16:00 266.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.5 F 29.1
2:00 271.6 F 28.8
3:00 272.5 F 29.1
4:00 271.5 F 28.4

Min 271.4 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 206.5 F 0.0
10:00 201.9 F 0.0
15:00 200.0 F 0.0
20:00 198.7 F 0.0
25:00 196.3 F 0.0
30:00 193.6 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 17:35
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:56:43
CYCLE COMPLETE

01/20/2022 12:02 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____