
Midmark M11 - v1.0.0
Total Cycles: 4192

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2024 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 71.3 F 0.2
4:00 84.7 F 1.0
6:00 113.5 F 2.9
8:00 144.0 F 5.3
10:00 168.7 F 8.1
12:00 181.9 F 11.2
14:00 207.3 F 4.0
16:00 212.9 F 5.1
18:00 219.3 F 8.2
20:00 229.5 F 11.8
22:00 230.5 F 6.9
24:00 239.7 F 11.7
26:00 237.0 F 9.1
28:00 249.3 F 14.8
30:00 259.5 F 20.5
32:00 269.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.4 F 29.3
2:00 272.1 F 28.8
3:00 272.0 F 29.1
4:00 272.6 F 29.1

Min 271.4 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.4 F 0.0
5:00 239.5 F 0.0
10:00 216.4 F 0.0
15:00 207.6 F 0.0
20:00 204.1 F 0.0
25:00 203.2 F 0.0
30:00 199.8 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 33:48
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:12:21
CYCLE COMPLETE

01/11/2024 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____