
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004632
9:26 02/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:23
STERILIZING 3:10
136.0 C 220kPa 3:10
136.3 C 220kPa 3:41
136.4 C 223kPa 4:11
136.2 C 220kPa 4:41
136.0 C 219kPa 5:11
136.1 C 221kPa 5:41
136.2 C 221kPa 6:11
136.4 C 224kPa 6:41
136.4 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:28
CYCLE COMPLETE 9:46

Digital Signature #
D20D06501C994C7B
