
Midmark M11 - v1.0.0
Total Cycles: 4973

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2024 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.6 F 0.0
2:00 160.7 F 0.8
4:00 183.4 F 4.6
6:00 210.5 F 8.9
8:00 227.6 F 9.8
10:00 226.7 F 8.7
12:00 239.5 F 11.4
14:00 241.6 F 11.6
16:00 239.5 F 10.2
18:00 252.5 F 17.0
20:00 264.2 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.7 F 29.1
2:00 271.7 F 29.1
3:00 272.1 F 28.9
4:00 272.1 F 29.3

Min 271.2 F 28.6
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 225.4 F 0.0
10:00 221.8 F 0.0
15:00 218.1 F 0.0
20:00 228.3 F 0.0
25:00 226.7 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:53
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

10/07/2024 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____