
Midmark M11 - v1.0.0
Total Cycles: 3997

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2023 03:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.0 F 0.0
2:00 90.6 F 0.7
4:00 114.8 F 2.6
6:00 144.5 F 5.6
8:00 191.3 F 9.2
10:00 200.2 F 13.0
12:00 217.3 F 5.0
14:00 229.0 F 8.5
16:00 236.8 F 12.6
18:00 234.0 F 8.3
20:00 245.3 F 13.5
22:00 242.7 F 11.4
24:00 254.6 F 17.4
26:00 265.4 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.1
1:00 272.5 F 28.8
2:00 271.4 F 28.0
3:00 272.4 F 28.5
4:00 271.5 F 28.2

Min 271.4 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 215.1 F 0.0
10:00 231.6 F 0.0
15:00 238.9 F 0.0
20:00 227.3 F 0.0
25:00 229.7 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 27:38
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:06:02
CYCLE COMPLETE

11/08/2023 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____