
Midmark M11 - v1.0.5
Total Cycles: 614

Name: RITTER2
Sterilizer ID: V2313228
12/08/2021 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 175.2 F 3.9
4:00 197.3 F 9.1
6:00 217.2 F 7.1
8:00 232.3 F 9.7
10:00 235.4 F 8.9
12:00 244.1 F 12.3
14:00 245.7 F 13.4
16:00 259.0 F 20.6
18:00 270.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.6 F 29.2
2:00 271.9 F 29.4
3:00 272.1 F 29.0
4:00 271.8 F 29.4

Min 271.3 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 238.0 F 0.0
10:00 229.3 F 0.0
15:00 198.8 F 0.0
20:00 188.0 F 0.0
25:00 183.3 F 0.0
30:00 179.2 F 0.0

Drying Complete

FILLING: 1:48
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 00:57:34
CYCLE COMPLETE

12/08/2021 03:26 PM
ACCEPTED: X
REJECTED:
Operator: U02