

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1528

Name: RITTER 1  
Sterilizer ID: V2178198  
01/14/2022 04:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.8 F 0.0  
2:00 78.5 F 0.0  
4:00 119.2 F 0.2  
6:00 127.1 F 2.4  
8:00 152.4 F 5.1  
10:00 173.3 F 8.5  
12:00 190.6 F 11.9  
14:00 205.0 F 2.7  
16:00 215.0 F 6.0  
18:00 224.9 F 9.4  
20:00 234.9 F 13.3  
22:00 234.7 F 9.0  
24:00 242.0 F 11.3  
26:00 243.9 F 12.1  
28:00 255.1 F 17.9  
30:00 265.2 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.2 F 28.9  
2:00 272.0 F 28.4  
3:00 272.1 F 28.8  
4:00 272.2 F 28.6

Min 271.5 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 201.0 F 0.0  
10:00 201.5 F 0.0  
15:00 202.5 F 0.0  
20:00 213.3 F 0.0  
25:00 245.4 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:50  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:09:59  
CYCLE COMPLETE

01/14/2022 05:28 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_