
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001942
8:28 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:56
STERILIZING 2:52
136.0 C 220kPa 2:52
136.4 C 225kPa 3:22
136.1 C 221kPa 3:52
136.1 C 222kPa 4:22
136.3 C 224kPa 4:52
136.0 C 222kPa 5:22
136.4 C 225kPa 5:52
136.5 C 225kPa 6:22
136.5 C 225kPa 6:22
VENTING 6:22
AIR DRYING 7:13
CYCLE COMPLETE 11:45

Digital Signature #
DCC8DA4A46B41B3C
