
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007900
7:45 01/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:03
STERILIZING 4:00
136.0 C 220kPa 4:00
136.2 C 221kPa 4:30
136.3 C 221kPa 5:00
136.2 C 221kPa 5:30
136.4 C 222kPa 6:00
136.4 C 221kPa 6:30
136.4 C 223kPa 7:00
136.4 C 219kPa 7:30
136.4 C 219kPa 7:30
VENTING 7:30
AIR DRYING 8:23
CYCLE COMPLETE 38:23

Digital Signature #
A6EF56FFF2D59550
