
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000483
9:57 18/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 3:28
CYCLE FAULT 15
NOT STERILE
END TIME 3:28

Digital Signature #
9649ACCEEE5EE61D
