
Midmark M11 - v1.0.5
Total Cycles: 831

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 04:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.3 F 0.0
2:00 176.8 F 3.6
4:00 199.8 F 8.8
6:00 221.9 F 8.7
8:00 233.2 F 10.1
10:00 234.0 F 8.3
12:00 241.5 F 11.1
14:00 247.2 F 14.1
16:00 260.1 F 21.1
18:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.6 F 29.3
2:00 271.3 F 28.6
3:00 272.4 F 29.5
4:00 271.9 F 28.8

Min 271.3 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 236.9 F 0.0
10:00 229.8 F 0.0
15:00 230.8 F 0.0
20:00 203.6 F 0.0
25:00 185.3 F 0.0
30:00 181.2 F 0.0
Drying Complete

FILLING: 2:46
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

02/23/2022 05:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____