
Midmark M11 - v1.0.5
Total Cycles: 3623

Name: RITTER2
Sterilizer ID: V2313228
05/15/2024 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.0 F 0.0
2:00 166.3 F 3.2
4:00 187.7 F 7.2
6:00 205.9 F 11.6
8:00 218.4 F 6.1
10:00 230.1 F 11.0
12:00 233.1 F 8.3
14:00 241.6 F 11.4
16:00 245.2 F 13.4
18:00 257.4 F 19.8
20:00 267.7 F 26.6
22:00 271.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.0 F 29.5
2:00 272.9 F 29.8
3:00 271.5 F 29.7
4:00 271.8 F 29.5

Min 271.2 F 29.3
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.6 F 0.0
5:00 234.1 F 0.0
10:00 195.0 F 0.0
15:00 192.1 F 0.0
20:00 187.7 F 0.0
25:00 184.6 F 0.0
30:00 183.2 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

05/15/2024 04:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____