
Midmark M11 - v1.0.0
Total Cycles: 5187

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2024 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.6 F 0.0
2:00 162.4 F 0.0
4:00 177.7 F 0.2
6:00 201.6 F 1.7
8:00 214.1 F 5.5
10:00 220.9 F 9.4
12:00 231.3 F 12.3
14:00 235.3 F 9.4
16:00 238.7 F 10.1
18:00 244.7 F 12.6
20:00 242.8 F 11.3
22:00 255.1 F 17.5
24:00 265.5 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.8
1:00 272.3 F 28.5
2:00 272.3 F 28.3
3:00 271.8 F 27.8
4:00 271.5 F 27.9

Min 271.4 F 27.7
Max 272.5 F 28.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 206.7 F 0.0
10:00 215.1 F 0.0
15:00 224.2 F 0.0
20:00 224.5 F 0.0
25:00 222.3 F 0.0
30:00 232.0 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 26:00
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

12/13/2024 01:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____