
Midmark M11 - v1.0.0
Total Cycles: 5172

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2024 11:01 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 81.6 F 0.0
2:00 105.2 F 0.0
4:00 142.8 F 0.1
6:00 175.0 F 0.3
8:00 199.2 F 0.8
10:00 217.6 F 3.5
12:00 227.2 F 6.6
14:00 235.8 F 10.2
16:00 236.4 F 9.1
18:00 242.3 F 11.5
20:00 238.3 F 9.3
22:00 236.3 F 8.0
24:00 251.9 F 15.9
26:00 265.2 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.6 F 28.4
2:00 272.2 F 28.4
3:00 272.4 F 28.2

Min 271.6 F 27.7
Max 273.0 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.7 F 0.0
Dry Cycle Aborted

FILLING: 0:36
HEATING: 27:39
STERILIZING: 3:00
VENTING: 2:33
DRYING: 1:02
TOTAL CYCLE: 00:34:50

12/09/2024 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____