
Midmark M11 - v1.0.5
Total Cycles: 4301

Name: RITTER2
Sterilizer ID: V2313228
01/27/2025 05:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.8 F 0.0
2:00 122.8 F 0.6
4:00 134.1 F 3.0
6:00 167.1 F 6.3
8:00 187.9 F 9.7
10:00 204.7 F 13.3
12:00 217.8 F 6.8
14:00 228.6 F 11.0
16:00 227.5 F 7.6
18:00 239.2 F 12.9
20:00 239.4 F 11.5
22:00 250.6 F 17.2
24:00 260.8 F 23.0
26:00 271.2 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 30.4
1:00 272.1 F 30.5
2:00 272.4 F 30.7
3:00 272.2 F 30.5
4:00 272.1 F 30.4

Min 271.9 F 30.3
Max 272.5 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 189.4 F 0.0
10:00 201.6 F 0.0
15:00 198.0 F 0.0
20:00 190.8 F 0.0
25:00 188.2 F 0.0
30:00 183.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 26:29
STERILIZING: 4:00
VENTING: 3:60
DRYING: 30:00
TOTAL CYCLE: 01:05:30
CYCLE COMPLETE

01/27/2025 06:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____