
Midmark M11 - v1.0.0
Total Cycles: 4152

Name: RITTER 1
Sterilizer ID: V2178198
12/22/2023 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 183.4 F 2.0
4:00 183.2 F 5.5
6:00 201.2 F 9.6
8:00 217.2 F 8.8
10:00 222.2 F 7.6
12:00 234.1 F 12.1
14:00 236.1 F 8.7
16:00 241.8 F 11.1
18:00 245.6 F 12.7
20:00 256.9 F 18.6
22:00 266.8 F 24.9
24:00 272.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.4
1:00 272.1 F 28.6
2:00 272.2 F 28.8
3:00 272.2 F 28.9
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 239.9 F 0.0
10:00 226.4 F 0.0
15:00 225.0 F 0.0
20:00 231.1 F 0.0
25:00 224.2 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:13
CYCLE COMPLETE

12/22/2023 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____