
Midmark M11 - v1.0.0
Total Cycles: 3972

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2023 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.8 F 0.0
2:00 151.7 F 2.7
4:00 168.4 F 6.3
6:00 200.7 F 11.3
8:00 222.3 F 5.0
10:00 229.5 F 10.0
12:00 233.8 F 7.8
14:00 243.6 F 12.4
16:00 241.2 F 10.7
18:00 254.2 F 17.2
20:00 265.7 F 24.1
22:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.5
2:00 272.5 F 29.0
3:00 271.5 F 28.4
4:00 272.4 F 28.7

Min 271.3 F 28.3
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.1 F 0.0
5:00 177.7 F 0.0
10:00 216.3 F 0.0
15:00 238.5 F 0.0
20:00 210.2 F 0.0
25:00 199.2 F 0.0
30:00 194.9 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

11/02/2023 11:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____