

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3779

Name: RITTER 1  
Sterilizer ID: V2178198  
09/15/2023 09:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.0 F 0.0  
2:00 162.3 F 0.6  
4:00 176.7 F 3.5  
6:00 198.7 F 7.4  
8:00 213.5 F 11.9  
10:00 227.2 F 6.6  
12:00 237.5 F 11.5  
14:00 236.0 F 8.5  
16:00 242.1 F 10.8  
18:00 246.7 F 13.3  
20:00 258.6 F 19.8  
22:00 269.3 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.1  
1:00 272.2 F 28.5  
2:00 271.6 F 28.4  
3:00 271.9 F 28.4  
4:00 271.4 F 28.4

Min 271.0 F 27.9  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.6 F 0.0  
5:00 203.0 F 0.0  
10:00 194.1 F 0.0  
15:00 185.1 F 0.0  
20:00 183.3 F 0.0  
25:00 183.5 F 0.0  
30:00 182.2 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 23:03  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:01:53  
CYCLE COMPLETE

09/15/2023 10:21 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_