
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001080
9:23 17/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:35
STERILIZING 3:37
136.0 C 220kPa 3:37
136.4 C 226kPa 4:07
136.4 C 226kPa 4:37
136.3 C 223kPa 5:07
136.4 C 223kPa 5:37
136.3 C 224kPa 6:07
136.2 C 223kPa 6:37
136.2 C 223kPa 7:07
136.2 C 223kPa 7:07
VENTING 7:07
AIR DRYING 7:58
CYCLE COMPLETE 22:58

Digital Signature #
D9C3EE377D63198A
