
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001478
14:13 10/06/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 5:29
STERILIZING 7:02
136.0 C 220kPa 7:02
136.2 C 221kPa 7:32
136.5 C 224kPa 8:02
136.4 C 224kPa 8:32
136.5 C 220kPa 9:02
136.3 C 221kPa 9:33
136.3 C 221kPa 10:03
136.2 C 222kPa 10:32
VENTING 10:32
AIR DRYING 11:35
CYCLE COMPLETE 41:36

Digital Signature #
3AD2ABBB282A2DBB
