
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002928
11:03 11/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:44
STERILIZING 4:51
136.0 C 220kPa 4:51
136.3 C 223kPa 5:21
136.3 C 223kPa 5:51
136.3 C 222kPa 6:21
136.4 C 225kPa 6:51
136.4 C 223kPa 7:21
136.4 C 223kPa 7:51
136.4 C 225kPa 8:21
136.4 C 225kPa 8:21
VENTING 8:21
AIR DRYING 9:19
CYCLE COMPLETE 9:20

Digital Signature #
7D84A9B95F3419AB
