

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2968

Name: RITTER 1  
Sterilizer ID: V2178198  
03/03/2023 10:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.3 F 0.0  
2:00 87.2 F 0.0  
4:00 100.9 F 0.0  
6:00 130.0 F 0.0  
8:00 173.7 F 2.2  
10:00 198.2 F 5.5  
12:00 220.3 F 9.8  
14:00 230.2 F 9.1  
16:00 228.4 F 7.6  
18:00 238.9 F 12.1  
20:00 236.7 F 8.9  
22:00 240.5 F 9.8  
24:00 248.0 F 13.8  
26:00 260.7 F 20.8  
28:00 272.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 271.8 F 28.2  
2:00 272.3 F 28.5  
3:00 272.7 F 28.7  
4:00 272.7 F 28.6

Min 271.6 F 27.7  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 238.1 F 0.0  
Dry Cycle Aborted

FILLING: 0:59  
HEATING: 28:21  
STERILIZING: 4:00  
VENTING: 2:48  
DRYING: 0:52  
TOTAL CYCLE: 00:36:60

03/03/2023 10:38 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_