
Midmark M11 - v1.0.5
Total Cycles: 1695

Name: RITTER2
Sterilizer ID: V2313228
11/08/2022 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.3 F 0.0
2:00 82.1 F 0.0
4:00 112.5 F 0.0
6:00 124.6 F 0.1
8:00 160.6 F 0.9
10:00 191.5 F 1.1
12:00 208.7 F 4.0
14:00 219.2 F 7.5
16:00 229.5 F 11.2
18:00 230.7 F 6.7
20:00 240.8 F 11.9
22:00 238.9 F 9.8
24:00 238.4 F 9.0
26:00 250.0 F 15.1
28:00 261.4 F 21.7
30:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.2 F 28.6
2:00 272.7 F 29.2
3:00 271.6 F 28.4
4:00 272.4 F 28.8

Min 271.6 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 204.4 F 0.0
10:00 221.4 F 0.0
15:00 224.0 F 0.0
20:00 224.2 F 0.0
25:00 226.3 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:26
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:08:19
CYCLE COMPLETE

11/08/2022 09:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____