
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005336
11:31 10/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:00
STERILIZING 3:05
136.5 C 220kPa 3:05
136.4 C 225kPa 3:35
136.2 C 222kPa 4:05
136.3 C 224kPa 4:35
136.4 C 225kPa 5:05
136.1 C 219kPa 5:35
136.4 C 222kPa 6:05
136.3 C 222kPa 6:35
136.3 C 222kPa 6:35
VENTING 6:35
AIR DRYING 7:23
CYCLE COMPLETE 12:03

Digital Signature #
560B3DB4416B9E34
