

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001976  
10:45 26/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:10  
PRESSURIZING 2:15  
STERILIZING 3:12  
136.0 C 220kPa 3:12  
135.8 C 221kPa 3:42  
136.4 C 225kPa 4:12  
136.3 C 224kPa 4:42  
136.4 C 223kPa 5:12  
136.1 C 222kPa 5:42  
136.4 C 225kPa 6:12  
136.4 C 226kPa 6:42  
136.4 C 226kPa 6:42  
VENTING 6:42  
AIR DRYING 7:32  
CYCLE COMPLETE 7:51  
-----

Digital Signature #  
EEEB2E7FF8A1A2C4  
-----