

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 004437  
13:20 26/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:13  
STERILIZING 4:15  
136.0 C 220kPa 4:15  
136.4 C 225kPa 4:45  
136.2 C 221kPa 5:15  
136.3 C 223kPa 5:45  
136.4 C 224kPa 6:15  
136.3 C 225kPa 6:45  
136.4 C 224kPa 7:15  
136.4 C 223kPa 7:45  
136.4 C 223kPa 7:45  
VENTING 7:45  
AIR DRYING 8:38  
CYCLE COMPLETE 9:01

-----  
Digital Signature #  
9DD8EE60FF3E951F  
-----