

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001121  
12:33 15/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:36  
PRESSURIZING 3:43  
STERILIZING 5:32  
136.0 C 220kPa 5:32  
136.3 C 223kPa 6:02  
136.3 C 223kPa 6:32  
136.4 C 224kPa 7:02  
136.2 C 222kPa 7:32  
136.0 C 223kPa 8:02  
136.3 C 223kPa 8:32  
136.3 C 223kPa 9:02  
136.3 C 223kPa 9:02  
VENTING 9:02  
AIR DRYING 10:05  
CYCLE COMPLETE 25:05  
-----

Digital Signature #  
5A428B90479E114F  
-----