

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3922

Name: RITTER2  
Sterilizer ID: V2313228  
09/10/2024 10:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.5 F 0.0  
2:00 123.9 F 0.1  
4:00 128.2 F 0.3  
6:00 159.1 F 3.2  
8:00 189.7 F 6.6  
10:00 207.4 F 10.9  
12:00 218.6 F 5.8  
14:00 227.6 F 9.4  
16:00 236.1 F 11.0  
18:00 239.1 F 11.4  
20:00 238.0 F 9.9  
22:00 250.1 F 15.9  
24:00 261.0 F 22.1  
26:00 270.9 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.4 F 29.7  
2:00 271.4 F 29.4  
3:00 272.4 F 29.7  
4:00 271.2 F 29.2

Min 271.2 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 206.1 F 0.0  
10:00 218.4 F 0.0  
15:00 199.6 F 0.0  
20:00 194.2 F 0.0  
25:00 202.9 F 0.0  
30:00 209.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:47  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:05:15  
CYCLE COMPLETE

09/10/2024 11:17 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_